Low-cost Active Electrode Improves the Resolution in Biopotential Recordings

Alexander C. MettingVanRijn, Anthony P. Kuiper, Taco E. Dankers and Cees A. Grimbergen.

Medical Physics Department, University of Amsterdam, Meibergdreef 15
1105 AZ Amsterdam, The Netherlands, e-mail: metting@amc.uva.nl

Abstract- The accepted models of the origin of interference in body surface potential recordings predict that it would be advantageous to perform signal amplification on the electrode. A suitable circuit for a miniature biopotential amplifier to be manufactured in thick-film technology is described. A prototype series of active electrodes was produced and tested. The high immunity to interference resulted in ECG and EEG recordings of a very high quality.

I. INTRODUCTION

The usual transducers in biopotential measurements (electrodes) are passive metal discs with a connecting wire to the electronic circuitry. This setup has the advantage of mechanical simplicity. However, from an engineering point of view it is unattractive to have a significant length of connection wire between a high impedance signal source (the electrode) and the measurement circuitry. Instead, the ideal solution would be to perform signal amplification as close to the transducers as possible.

There are indeed a number of practical problems with the current passive electrodes. Unshielded electrode wires are usually the major source of power supply line (50 Hz) interference; the mechanism has been described in a previous paper [1]. A solution can be formed by using shielded electrode wires and although good suppression of power supply line interference can be achieved with this method, widespread clinical use is hampered by the costs and by the delicacy of the low-noise shielded wires and the coaxial connectors.

Another problem with the passive electrodes in modern biopotential measurement systems originates from the current trend to combine low-noise analog circuitry and fast logic in the front-end. It was argued in an earlier paper that optimal instrumentation in terms of interference rejection and patient-safety should consist of a small battery powered front-end with a fiber-optic connection to the signal-processing hardware (usually a PC) [2]. With currently available components, the best solution is to perform the analog to digital conversion in the isolated front-end and to use a digital transmission format for the fiber-optic transmission link. A consequence of such a setup is, however, that special attention has to be paid to the problem of interference from the digital circuitry coupled into the sensitive analog input circuitry.

The two described problems could in principle be solved if the analog signal processing would be performed at the electrode. Such an active transducer would form a signal

source with a low output impedance which would eliminate the need for shielded wires. In addition, proper shielding of the digital circuitry with respect to the analog input amplifiers would be feasible.

II. METHODS

Other instrumentation groups have developed active electrodes in the past. The electronics of these transducers always consisted of a buffer amplifier. Although this setup solves the problem of cable interference and can be designed to need only a two lead connection wire [3], there remains an important disadvantage. Because the signal is not amplified there is still needed an extensive low-noise amplifier stage in the front-end with its described problems with respect to HF interference. Therefore, we chose a different approach by attempting to perform all analog signal processing, e.g. amplification and DC rejection, at the electrode and to retain the usual configuration of amplifying the biopotential signal with respect to a reference signal [4].

The main design goals were to keep the size and costs of the electrode comparable to a standard EEG electrode. The cost factor eliminated the possibility of a custom made integrated circuit. We found a good alternative in combining thick-film technology with an amplifier design developed by our group to be used with a minimal number of parts [4]; the circuit diagram is shown in Fig. 1.

The electronic components are mounted at one side of a ceramic substrate with a diameter of 11 mm, while the other side is covered with a printed gold layer to form the electrode surface. The electronics are sealed with epoxy. The electrode is equipped with a 4 lead ribbon wire of 1 meter length with a small low-cost crimp connector. No expensive components or production processes have been used.

III. RESULTS

The active electrodes were tested in combination with a fiber coupled 32-channel front-end. The front-end consisted of a 16-bit ADC for each channel, an optical fiber transmission circuit, timing logic and a NiMH battery pack, housed in a compact box positioned close to the patient. The system was interfaced to a standard PC by a commercially available digital I/O card.

In order to test the new setup, EEG and ECG signals were recorded with the electrodes applied to the unprepared skin with a typical electrode paste (Elefix, Nihon Kohden Corporation). The measurements were performed in a typical

laboratory room full of electrical equipment. All measurements showed an exceptionally low noise level and a complete absence of 50 Hz interference. One of the EEG measurements is shown in Fig. 2. Note that the shown signals are raw data, no data processing other than bandwidth limiting has been performed.

IV. DISCUSSION

Although the signal quality of the active electrodes is satisfactory, there are some practical problems to be solved before the electrodes can be regarded to be an alternative for the commonly used passive electrodes.

The cost of this new technology will be a decisive factor determining the general acceptance of this new type of electrodes. Production costs were kept minimal by using only well established thick-film production techniques and by designing the electronics around easily available standard components (1 dual opamp in naked chip version, 8 SMD metal-film resistors and 1 SMD capacitor). The projected production costs for large series is in the order of US\$ 10.

Some practical details need further improving. The current used ribbon cable is cheap and easily mounted with a connector but not really flexible enough for day-to-day clinical use. Another point of further research is the combination of the chosen electrode material and the various electrode pastes which at the moment is the limiting factor for the signal quality.

V. CONCLUSIONS

It was shown that a high signal quality can be achieved with the developed active electrodes on the unprepared skin. We expect that using this method of concentrating the analog circuitry on the electrode will be the most economical way of producing high-quality biopotential recording systems.

ACKNOWLEDGEMENTS

This work was supported by the Technology Foundation 'STW under grant No. AGN44.3416.

REFERENCES

- [1] A. C. MettingVanRijn, A. Peper and C. A. Grimbergen, "High-quality recording of bioelectric events. I: interference reduction, theory and practice," *Med. & Biol. Eng. & Comput.*, vol. 28, pp. 389-397, 1990.
- [2] A. C. MettingVanRijn, A. P. Kuiper, A. C. Linnenbank and C. A. Grimbergen, "Patient isolation in multichannel bioelectric recordings by digital transmission through a single optical fiber," *IEEE Trans. Biomed. Eng.*, vol. 40, pp. 302-308, 1993.
- [3] F. Z. Padmanidina, J. J. Veerhoek, G. J. A. VanDijk and J. H. Huijsing, "Microelectronic skin electrode," *Sensors and Actuators*, vol. B1, pp. 491-494, 1990.
- [4] A. C. MettingVanRijn, A. Peper and C. A. Grimbergen, "Amplifiers for bioelectric events: a design with a minimal number of parts", *Med. & Biol. Eng. & Comput.*, vol. 32, pp. 305-310, 1994.

Fig. 1 The circuit diagram and transfer function of the active electrode. The stopped integrator in the feedback loop reduces the gain for low frequencies in order to keep the relatively large electrode offset voltages within the dynamic range of a 16 bit ADC.

Fig. 2 An EEG signal measured with the active electrodes on the unprepared skin. Note the low noise level of approx 3 $\mu\text{V}_{\text{pk-pk}}$ and the absence of 50 Hz interference.