



Department of Environmental Services

Environmental Laboratory Report

Maggie Brooks
County Executive

Michael J. Garland, P.E.
Director

Folder Number: 124743

Page 1 of 1

Sample Origin: RIT

Report Date: 03/04/2011

Project Name: Special Sample

Sampling Event: RIT

Sample Type: 1G

Folder Comments:

Accepted By: Brown, G

Analyte	Results	Units	Method	Test Date/Time	Analyst	Status	Comments
---------	---------	-------	--------	----------------	---------	--------	----------

Sample ID:	410608	Received at Lab:	02/09/2011 @ 1505	Comments:				
Sample Site:	RIT- Crumb Rubber	Collector:	Burleigh, Bettina	Sample Date:	02/08/2011	Sample Time:	1738	
Arsenic, Total	< 0.00100	mg/L	SM 18-19, 21 3113B	02/25/2011 @	Brown, G	Approved		
COMMENT:	Estimated Value: Value is between the MDL and the lowest standard. Post Digested Spike biased high							
Cadmium, Total	< 0.000100	mg/L	SM 18-19, 21 3113B	02/23/2011 @	Brown, G	Approved		
COMMENT:	Post Digested Spike biased high							
Chromium, Total	< 0.00100	mg/L	SM 18-19 3113B	02/18/2011 @	Brown, G	Approved		
COMMENT:	Estimated Value; Value is between the MDL and the low standard							
Copper, Total	< 0.00500	mg/L	SM 18-19 3113B	03/01/2011 @	Brown, G	Approved		
Lead, Total	< 0.00100	mg/L	SM 18-19, 21 3113B	02/18/2011 @	Brown, G	Approved		
COMMENT:	Post Digested Spike biased high							
Manganese, Total	0.174	mg/L	SM 18-19, 21 3113B	03/02/2011 @	Brown, G	Approved		
Nickel, Total	0.00260	mg/L	SM 18-19 3113B	03/01/2011 @	Brown, G	Approved		
COMMENT:	Estimated Value; Value is between the MDL and the lowest standard Post Digested Spike biased low							
Zinc, Total	4.23	mg/L	SM 18-1213111B (99)	03/03/2011 @	Goodrich, C	Approved		

(1) Not currently certified for this analyte.

(2) Currently no certification is available for this analyte

(3) NYSDOH ELAP Certified only

These results relate only to the samples received by the laboratory for analysis

NYSDOH ELAP # 10373

EPA Lab Code NY01266



Approved By:

Scott Gabel, QA/QC