

14026 WEEK 9 DEMO

Mike Allocco
Soham Chakraborty
Leslie Havens
Danielle Koch
Andrew Miller
Kristeen Yee
Stephanie Zambito

AGENDA

- Team Status
- Pressure Sensors & Board
- Orifice Plates & Trachea Resistance
- Actuation
- Leak Tests
- Flow Meter Validation
- Lung Compliance
- Maximum Lung Volume
- Demonstrate prototype can be hooked up to read and store values
- Set Expectations

TEAM STATUS

	Status	Comments	Lead
Week 3-5 Planned Tasks			
Demonstrate LabView can be used to log in values read by the pressure sensor	Green		Mike
Test that pressure sensors are	Green	Done	Leslie
Present a soldered PCB board with established connections to pressure sensors	Green		Leslie
Test A5: Pressure Measurement Capability (test pressure sensors are accurate)	Red	We have a functional relationship but not a significant amount of points.	Andrew & Soham



TEAM STATUS

Week 6-8 Planned Tasks			
Build the remaining orifice plates		Functional but need slight adjustments.	Andrew
Build actuation setup		Done	Danielle
Program actuation		Done	Mike
Install and test DAQ Board		Done	Soham
Finish updates to LabView program		Need to add the solenoid valve	Mike
Demonstrate four modes can be read and stored by the tester		Need final assembly completed	All
Test A1: Leak Test Configuration 1		Done	Kristeen & Stephanie & Leslie
Test A2: Flow Meter Validation		Customer during last demo said assuming flow meter was correct is fine. If time later will try to come back and put constant flow through it.	Mike & Danielle
Test A3: Lung Compliance		Done	Leslie & Danielle & Kristeen
Test A4: Leak Test Configuration 2		Done	Andrew & Soham
Test A6: Resistance of Trachea & Lung (airway resistance analysis)		Done	Andrew
Test A7: Leak Test Configuration 3		Need final assembly completed	Kristeen & Stephanie
Test A8: Maximum Lung Volume		Done	Danielle & Mike


PRESSURE SENSORS & BOARD

	Identifying & Selecting Problem PSP 1	Analyzing Problem PSP 2	Generating Potential Solutions PSP 3	Selecting & Planning Solution PSP 4	Implementing Solution PSP 5	Evaluating Solution PSP 6
Rating	R1	R2	R3	Y4	Y5	G6
CRITICAL	The pressure sensor is not reading a voltage change on the o-scope.	1) Pressure sensor is defective. 2) Team does not know what they are doing.	1) Seek expert advice to confirm that team has it hooked it up right. 2) We could buy additional pressure sensors. 3) Look for different pressure sensors. 4) Switch to a digital manometer.	Seek expert advice to confirm the team has hooked it up right.	Went over testing setup with Dr Slack. Suggested that we get a breakout board to check that pins are not touching. Having Dr Fuller build us a breakout board. Will resume testing to evaluate the solution.	Using the breakout boards we found that the newly purchased pressure sensors did show a voltage change for change in pressure.
	Op-Amp started smoking when power was applied.	1) Traces are incorrect, input voltage pin was grounded. 2) Op-Amp orientation was wrong. 3) Soldering the pressure sensor pins could have caused a short. 4) Op-Amp is defective.	1) Continuity testing 2) Analyze all chip pin outs and compare with PCB layout. 3) Test op-amp directly by applying power and observing voltage with o-scope.	Analyze all chip pin outs and compare with PCB layout. Then, perform continuity testing.	Looked at the connections online versus the PCB layout. Team found that input voltage was grounded. Team will create a new board layout and build on a breadboard.	The op-amp works on the breadboard. This was tested by hooking it up to the o-scope.

PRESSURE SENSORS WORKING




NEW LAYOUT DESIGN




LABVIEW WITH PRESSURE SENSORS

- Labview measures voltage changes associated with increased pressure.



TESTING PRESSURE SENSORS


- Accuracy of pressure sensors
- Piezoelectric material used in sensors
- Must be calibrated, 97% linear relationship between volume and voltage



ORIFICE PLATES

	Identifying & Selecting Problem PSP 1	Analyzing Problem PSP 2	Generating Potential Solutions PSP 3	Selecting & Planning Solution PSP 4	Implementing Solution PSP 5	Evaluating Solution PSP 6
Rating	R1	R2	R3	Y4	Y5	G6
MAJOR	Having trouble getting in the queue to waterjet the orifice plates.	1) Could not get a hold of Mr Bonzo.	1) Build the orifice plate in the machine shop. Pros- plate will be built. Cons- it will not be as precise. 2) Continue to try to get a hold of Bonzo. Pros- will hopefully get orifice plate waterjetted. Cons- Need to move forward.	Continue to try to get a hold of Mr Bonzo	Found Mr Bonzo in his office after MSD meeting.	The team has an orifice plate built.
ORDINARY	Unexpected results from first round of orifice plate validation testing	1) Empirical correction coefficients were inaccurately used for this circumstance 2) Plate has incorrect diameter	1) Measure the orifice plate diameter and confirm or disprove its diameter is within tolerance of nominal 2) Run a second series of tests and determine/ refine correction coefficients to determine adjusted diameter.	Measure the orifice plate diameter. If diameter assumption is incorrect, run calculations to see if results match new theory calculation.	Measured the diameter of the orifice plate. Found that the diameter was not as assumed after cutting. Re-calculated the theoretical pressure drop across the plate.	Found that with the measured diameter, the theoretical calculations matched the initial testing.

AIRWAY RESISTANCE TEST SETUP



AIRWAY RESISTANCE RESULTS

R=5

Round 1 of Testing

For Nominal critical diameter of 4.955 mm										For actual critical diameter of 4.674 mm			
	Measured Q (averaged) [Lpm]	Experimental Values ΔP (cmH2O)					Expected Value	Percent Difference	Measured Q (averaged) [Lpm]	Expected Value (cmH2O)	Measured Value (cmH2O)	Percent Difference	
		Trial 1	Trial 2	Trial 3	Trial 4	Avg							
Upper Bound Flow Rate	26.08	5.00	5.20	5.00	5.00	5.05	2.17	132.718894	26.08	2.27	5.05	122.4669604	
Lower Bound Flow Rate	11.04	0.50	0.33	0.50	0.33	0.42	0.92	54.71014493	11.04	0.67	0.42	36.84210526	
Median Flow Rate*	18.50	2.25	2.00	2.00	2.00	2.06	1.55	33.06451613	18.50	1.87	2.06	10.21936865	
Equivalent Resistance value for median flow rate = 6.75													
Design Point	* This was the flow rate used for the design of the orifice plate(s): designed for 21LPM and actual median flow rate is approximately 18.5 LPM							This number is expected to be improved after the second round of testing with the improved test procedure. Then orifice diameter can be altered with a predictable outcome.					

Round 2 of testing

	Measured Q [Lpm]			Experimental Values ΔP (cmH2O)			Expected Values			Percent Difference		
	Trial 1	Trial 2	Trial 3	Trial 1	Trial 2	Trial 3	Trial 1	Trial 2	Trial 3	Trial 1	Trial 2	Trial 3
Upper Bound Flow Rate	25.5	25.6	26.05	4.5	4.5	4.75	3.55223219	3.580147464	3.707118254	26.6809082	25.6931	28.1319
Lower Bound Flow Rate	12.28	12.26	12.38	0.25	0.25	0.50	0.823792281	0.821111104	0.837263722	69.652544	69.5535	40.2817
Median Flow Rate*	18.6	18.65	18.6	2	2	2	1.889935023	1.900109621	1.889935023	5.823743983	5.25709	5.82374
Equivalent Resistances												
	Upper Bound Flow Rate		10.5882									
	Lower Bound Flow Rate		1.22349									
	Median Flow Rate*		6.45161									
	Average Upper / Lower		5.90586									


ACTUATION




SYRINGE SIZE VALIDATION

- Using a 10 mL syringe
 - A breath was not triggered by pulling on the syringe
- Using a 60 mL syringe
 - A breath was triggered by pulling on the syringe
- Conclusions
 - The team plans to test the range of negative pressure that the PEV can detect. This will then be compared to our own gasps of breath.

SYRINGE SIZE VALIDATION



LABVIEW PROGRAM



LEAK TESTS WITH DISH SOAP

Testing theory:



Connection Test:



LEAK TESTS CONFIGURATION 2

Test Theory:



Leak Test:



FLOW METER VALIDATION


- During the last demo, the customer was satisfied with just assuming the flow meter was correct to the manufacturing specs.
- Flow Measurements:
 - Measurement Range: 0 to 300.0 Std L/min
 - Accuracy: 2% of readings 0.05 Std L/min, whichever is greater, at standard conditions (21.1°C and 101.3 kPa)
 - Response Time: less than 4 msec, 63% of final value at full scale flow

LUNG COMPLIANCE




- Box sits on lung as lung inflates
- Weights are measured in grams
- Box weighs 65g

LUNG COMPLIANCE



MAXIMUM LUNG VOLUME

- The lung was slightly more than 1 L
- Team will not have to worry about the nominal value as limit to be avoided to protect instrumentation.



DEMOS OF PROTOTYPE

- Show that it can be hooked up to various PEV's

TEAM VISION WEEKS 9-11

At the end of MSD II phase 4 the team will have a demonstration of our project's progress. Our expectations for our demonstration are:

- Demonstrate that our prototype can collect and store data for the four modes of the MediResp IV.
- Complete test plans S1-S12
 - Be able to show the customer our test results for tests S1-S12.
 - Be prepared to describe how we tested each requirement.
 - Be prepared to show how we tested each requirement.
- Create a database to display reports of the data collected for each mode.
- Set expectations for the next phase review:
 - Complete all test plans
 - Prepare and run our Imagine RIT exhibit.
 - Complete our project poster and write-up.
 - Submit application for poster contest.
 - Begin project close out.

TEAM ROLES WEEKS 9-11

Week 9-11 Planned Tasks	
Test S1: Breath Rate	Danielle
Test S2: Inspiration Time	Stephanie
Test S3: Rise Time	Kristeen
Test S4: Tidal Volume	Leslie
Test S5: Inhale:Exhale Ratio	Soham
Test S6: Pressure Provided	Mike
Test S7: Weight	Leslie
Test S8: PEV-Human Connection	Soham
Test S9: Extrinsic Peak End-Expiratory	Andrew
Test S10: Mean Airway Pressure	Danielle
Test S11: Intrinsic Peak End-Expiratory Pressure (AKA Auto-PEEP)	Danielle
Create database to display reports of the data collected for each mode	Kristeen
Develop elevator speech for Imagine RIT	All

QUESTIONS?

