

Action Items					
Item #	Description	Responsible	Due Date	Close Date	Comments
A001	Consider Permanent Elastic in front	Noah & Tyler	11/11	IP	By DDR
A002	Refine foot-lift model with angles	Noah	11/11	IP	By DDR
A003	Consider not using a quick connect	Noah & Tyler	-	11/06	Gate Review
A004	BOM: add columns (vendor/shipping)	Geni	11/11	11/06	By DDR (long term)
A005	Look into downsizing electrical side	Megan	11/11	11/10	By DDR (long term)
A006	Add music to videos	Noah	-	11/18	(optional :)
A007	Color correlation on the M-opt plot	Noah	11/11	IP	Make EGDE clearer
A008	Engineering metric for foot attachment	Geni	11/11	11/18	Need to quantify comfort
A009	Ask Dr. C about compression sleeves	Geni & Jared	11/11	10/30, 11/11	Nazareth clinic
A010	Tweak/update sketched	Geni	11/11	IP	GAD models
A011	Color code the risk table with changes	Tyler	11/11	11/11	Expect likelihood changes
A012	Test strain of fishing line	Noah	11/11	11/13	Hang weights on the line
A013	Increase severity of fishing line risk	Tyler	11/11	11/04	See other strain risk
A014	Move pressure alert priority down	Megan	11/11	11/06	On prioritized task list

Issues					
Item #	Description	Responsible	Open Date	Close Date	Comments
I001	The AFO could be slippery with socks	Geni	10/23	IP	Relates to A009

Decisions				
Item #	Description	Contributing Individuals	Decision Date	Comments
D001	Testing comfort with different users	Geni - Tyler	10/23	This is an ongoing resolution
D002	Bring Upper-Lower attachment prototype	Geni - Tyler	10/23	Decided as goal for DDR