

Action Items					
Item #	Description	Responsible	Due Date	Close Date	Comments
A001	Color correlation on the M-opt plot	Noah	DDR	12/9	Make EGDE clearer
A002	Weigh the shoe for the foot-lift	Noah/Geni	DDR	12/8	Inconsistency in model
A003	Repeatability of the muscle tests	Tyler/Noah	DDR	11/25	At least three trials
A004	Modify/improve muscle attachment	Tyler	DDR	12/1	Dove tail/vertical support
A005	Pressure not flow in schematic	Megan	DDR	12/1	
A006	Pseudocode	Megan	DDR	12/1	
A007	Consider 3M rather than Velcro	Jared/Noah	DDR	12/7	For component housing
CM01	Comment in the customer requirements	Noah	DDR	11/25	FT2 “does not include spasticity”
CM02	Move hip angel data to private SVN	Jared	DDR	12/1	In ME Research
CM03	Develop a Working Model with sensitivity analysis	Noah/Tyler	MSD II wk2		In MechE labs
CM04	Formalize MSD II Test plans	Noah	DDR	12/7	As outlined in Agenda
CM05	Comment Engineering Requirements	Noah	DDR	11/25	“May not be able to reach
CM06	Begin IRB permissions	Geni	MSDII	11/25	Deadline flexible
CM07	Consider easy plantar flexion release	Tyler	MSDII	IP	Manual support too?

Issues					
Item #	Description	Responsible	Open Date	Close Date	Comments
I001	The AFO could be slippery with socks	Geni	10/23	IP	Relates to A009